

ULTRAHIGH FIELD NMR AND MRI: SCIENCE AT A CROSSROADS
WORKSHOP REPORT

Tatyana Polenova¹, Thomas F. Budinger²,
¹University of Delaware; ²University of California Berkeley

TABLE OF CONTENTS

I. INTRODUCTION AND CONTEXT

A. Present state of the art in magnetic resonance.....	3
B. 2015 Workshop “Ultrahigh Field NMR and MRI: Science at a Crossroads”.....	5
• Objectives.....	6
• Format.....	6
• Outcomes.....	6

II. OUTCOME: UHF NMR/MRI IN THE UNITED STATES - ENABLING TRANSFORMATIONAL SCIENCE

A. UHF MR: Transformational science drivers.....	7
B. UHF MR: New technologies and their development.....	13
C. UHF MR: Instrument deployment and funding mechanisms, shared facilities infrastructure.....	20

III. SUMMARY AND NEXT STEPS.....23

IV. REFERENCES CITED.....24

V. APPENDICES

A. Toward a roadmap on UHF MR science for the US.....	25
B. List of workshop participants.....	27
C. Workshop program and presentation titles.....	35

I. INTRODUCTION AND CONTEXT

A. Present state of the art in magnetic resonance

Magnetic resonance plays a central role in academic, industrial and medical research. *Nuclear magnetic resonance (NMR)* is widely used for characterizing the structure, chemistry and dynamic properties of new materials, chemicals and pharmaceuticals, in both the liquid and solid phases. NMR also provides detailed functional information on biological macromolecules and their assemblies, *in vitro*, in membranes and even in whole cells. *In vivo, NMR imaging (MRI) and spectroscopy (MRS)* are used for clinical diagnosis and prognosis of disease, for non-invasive studies of human physiology and metabolism in general, and for evaluating brain function, in particular. MRI/S is also a key technology for imaging small organisms at the cellular level, monitoring catalysis in chemical reactors and other scientific areas where non-destructive characterizations of structure and dynamics in complex systems are needed.

At the heart of all the MR methods are strong, stable and homogeneous magnets built from low-temperature superconductors (LTS), which are essential to these experiments. Further developments in NMR/MRI are hampered because the ultimate limit of the attainable field strengths of persistent LTS magnets has now been reached. Fortunately, recent breakthroughs in new high-temperature superconductors (HTS) and hybrid LTS/HTS magnet technologies promise to greatly increase the achievable field strength of NMR magnets and to decrease the operational complexity of high field human MRI infrastructures, thereby enabling new applications at the forefront of modern multidisciplinary research. What is required, however, is a mechanism for the deployment of “ultrahigh field” magnets (above 23.5 tesla for NMR and above 11.7 tesla for MRI) into shared instrumentation facilities across the nation. The US is presently lacking in such a mechanism, and the resulting lack of infrastructure is hurting further progress in MR-based sciences in the US. These deficiencies have tremendously adverse impacts on our ability to engage in cutting-edge science in several areas of strategic national importance –including several that are main foci of NSF, DOE and NIH interests (*vide infra*). Furthermore, the lack of advanced infrastructure in ultrahigh field NMR/MRI in the nation impacts negatively our ability to train the next generation of scientists who will be the future leaders of the US research enterprise.

The NMR equipment market has grown significantly in the years since Nobel prizes were awarded to NMR scientists Richard Ernst (1991), Kurt Wüthrich (2002), Paul Lauterbur and Peter Mansfield (2003). Further progress in NMR and MRI/MRS, however, is confronting limitations inherent to the technology that has been used since the 1970s to build the ultra-stable, ultra-high-resolution magnets – with respect to both stability and homogeneity parameters at sub-ppm levels.¹ Current magnet technology relies on low-

¹The recent decision by Agilent to exit the NMR and MRI business was due to their declining market share and profitability in the face of strong investments in R&D by their competition, and reflects the importance of continued technological development to NMR and MRI science.

temperature superconductor (LTS) that, as illustrated in **Figure 1**, have reached an ultimate limit of 23.5 tesla (T) –corresponding to 1 GHz proton Larmor frequency. While very efficient at fields <12T, the cost of LTS-based magnets doubles per 5-10% increment above the 15 T level. Further efforts based on the *preconception* that only persistent LTS-based technologies can deliver the field strengths, stabilities and homogeneities needed to perform MR, will result in enormously large and expensive new magnets, whose siting would be limited to very few facilities. This cost issue has already hurt fundamental




Figure 1: Progress in LTS-based NMR fields, achieved over the last 44 years.

developments in magnetic resonance-based science in general, and the US MR-based community in particular. Currently, there are a total of fourteen NMR systems worldwide operating at magnetic field strengths of 22.3 – 23.5 T, of which seven are in Europe, four are in Asia and Australia, and only two are in the US. Two additional 23.5 T systems are undergoing installation in Europe (Bayreuth, Germany) and North America (Toronto, Canada). Seven more 25.8-28.2 T NMR instruments have been ordered by European institutions: one at 25.9 T (Zürich, Switzerland), and six at 28.2 T (Florence, Italy; Göttingen, Germany; Utrecht, The Netherlands; München, Germany; Lille, France; and Jülich, Germany). Additionally, three more orders are planned from institutions in Europe- for two 23.5 GHz and one 28.2 T systems. Constraints have also limited the availability of commercial human MRIs, which can be purchased up to only 7 T for clinical and research applications. Currently, the highest field magnets operating for human MRI studies (9.4 T) are located at the University of Chicago and the University of Minnesota (9.4 T and 10.5 T), with 11.7 T systems planned at NIH and Saclay (France), and a proposed system at 14 T in South Korea. Continued development of ever larger and more expensive high field magnet systems is unsustainable, yet scientific progress critically depends on access to higher fields.

Without exploiting new magnet technologies, the utility of many forms of magnetic resonance – including flagship techniques that have revolutionized chemical, materials, biological and medical sciences - cannot be fully realized. In 2013 two “grand challenges” were issued for which NMR and MRI are of fundamental relevance. One stemmed from the 2013 Report of the National Research Council “Magnetic Field Science in the United States: Current Status and Future Directions”. The report compiled by a panel of experts at the request of the National Science Foundation and the




Figure 2: Maximum current density of superconducting wires at 4.2 K as a function of applied magnetic field, contrasting the LTS (blue, purple) vs. the HTS (red) materials.

Department of Energy [1], called for (1) the US to regain its competitiveness in the field of high resolution NMR by the development and installation of spectrometers based on magnets that approach 30 T by 2020

(Figure 1), and (2) development of a human MRI/NMR system for brain research and physiology studies at 20 T. A second grand challenge relates to the BRAIN initiative, which calls for the development of new non-invasive methods for imaging and understanding how the healthy human brain functions and is organized, including the development of advanced MRI systems leading to higher sensitivity and better spatial resolution. Both of these challenges cannot be adequately addressed if constrained by the “only with persistent LTS-based magnets” preconception: LTS simply cannot cross the 24 T boundary to deliver the magnetic fields required for further progress in materials and biophysical NMR (Figure 2), and 20 T human-capable MRI systems relying on LTS technologies would lead to magnets weighing in excess of ≈400 tons. Moreover, the costs associated with developing these ultrahigh field magnets and the understandable reticence of commercial vendors to assume the associated risks, are prohibiting the development of new technologies that meet these challenges. Fortunately, recent progress in new conductor and magnet technology now presents the opportunity to overcome the limitations of LTS wires (*vide infra*).

B. 2015 Workshop “Ultrahigh Field NMR and MRI: Science at Crossroads”

Objectives. Having recognized an urgent need to move forward at this time to address the pressing needs of the scientific community that require new generation of ultrahigh field MR technologies and infrastructure in the US, a workshop “Ultrahigh Field NMR and MRI: Science at Crossroads” was initiated by the scientific community. The workshop, supported by the National Science Foundation, the Department of Energy, and the National Institutes of Health, took place on November 12-13, 2015, in Bethesda, MD, on the NIH campus.²

The specific objectives of the workshop were:

Objective 1. To establish the science drivers for which ultrahigh field magnetic resonance capabilities are essential, and which will lead to transformative applications in strategic areas of national interest.

Objective 2. To discuss the opportunities for the development of novel, cost-effective ultrahigh field NMR and MRI magnets on the basis of HTS and serial LTS/HTS technologies.

Objective 3. To develop a consensus among scientists and agencies regarding the optimal roadmap for the development of ultrahigh magnetic field technologies, and for enabling the acquisition and installation of such instrumentation in national and regional shared user facilities in the United States.

² Prior to this workshop a National Research Council report entitled “Current Status and Future Direction of High Magnetic Field Science in the United States” made observations and recommendations consistent with the findings and directions for future strategies concluded by the workshop participants (1),

Format. The meeting brought together scientists from academia and industry who develop and use ultrahigh field (UHF) technologies and applications, as well as Program Officers from NSF, DoE, and NIH (for the list of participants, see **Appendix B**). Both the science opportunities and the status of high magnetic field technologies were reviewed. The workshop was comprised of plenary and breakout sessions, featuring talks and round table brain-storming sessions, to discuss key technological issues and applications to a broad range of problems, and to establish the basis for developing a roadmap for the development of ultrahigh field technologies that will enable transformative science. The meeting was highly interactive with ample opportunities for everyone to actively participate. In preparation for the meeting, the participants worked in nine task forces, to identify the key issues to be discussed at the workshop. These reports are presented on the Workshop's website (<http://sites.udel.edu/uhf-nmr-workshop/>). A blueprint for the development of a roadmap that could enable UHF NMR/MRI science in the US is presented in **Appendix A**.

The workshop materials, including the presentation slides and abstracts, can be found on the workshop's website: <http://sites.udel.edu/uhf-nmr-workshop/>

Outcomes. The workshop's aim was to catalyze the development of a long-term UHF magnetic resonance science program in the United States. Major outcomes of the workshop included a summary of the new scientific frontiers that would be opened by the advent of UHF NMR and MRI, and a summary of the technological breakthroughs that would be needed to enable these. These summaries are presented in Section II. A strategy to make available to the broad scientific community such UHF facilities with the resources to address these crucial high field scientific drivers is underway. To further enable this vision we will seek feedback from the broader scientific community, encompassing both MR experts and non-MR scientists. Key sections of this final roadmap should include i) the transformational science that would arise from ultrahigh magnetic field MR; ii) the development of new technologies, including new HTS materials, that are needed to enable sustainable production and operation of UHF NMR and MRI instruments; iii) specific recommendations for the sustainable models of shared UHF MR centers in the US, including alternative concepts for collaboration between the federal agencies, national laboratories, and industry to enable advancement of technologies and support of proposed facilities. It is our intent to broadly disseminate this final roadmap document throughout the US scientific community and its various funding agencies (including but not limited to DoE, NIH, NSF, DoD), in the expectation of enabling the realization of its recommendations within the coming years.

II. OUTCOME OF THE UHF NMR/MRI WORKSHOP:

UHF NMR/MRI IN THE UNITED STATES - ENABLING TRANSFORMATIONAL SCIENCE

A. UHF MR: Transformational science drivers

A major focus of the meeting was, through presentations and roundtable/breakout discussions, to compile scientific directions of strategic national importance, which must rely on UHF MR to reach their realization. The participants concluded that the new science that will be enabled by the UHF technologies extends throughout wide swaths of physics, chemistry, materials science, biology, engineering, biogeochemistry, and biomedical sciences. These transformational science programs span multiple disciplines, with each identified direction being of strategic interest to at least two of the three Federal funding agencies that attended the Workshop. This conclusion is not surprising, given the central role of magnetic resonance in the molecular systems sciences. Magnetic resonance, in its spectroscopic (NMR) and imaging (MRI) modalities, is the only method that provides an integrated view of static structure, dynamics, and molecular mechanisms, in a non-destructive and site-specific fashion, often in a single set of experiments. The knowledge of the molecular structure and dynamics are gained at atomic resolution (NMR) or at micrometer resolution (MRI). MR techniques are capable of probing the broadest range of systems, many of which are intractable by other structural techniques. The main challenge in the molecular systems sciences is to connect the *in vitro* information to *in vivo* or *in situ* pictures, and magnetic resonance methods are uniquely positioned to provide this link. For this purpose, the highest resolution and the highest sensitivity are required across the board, necessitating the development and application of UHF technologies.

Key examples of scientific directions requiring UHF MR technologies and infrastructures are presented below. The workshop participants concluded that gains in sensitivity, resolution, and information content are highly nonlinear with the magnetic field strength, and ultrahigh fields will enable fundamental investigations of emergent phenomena not accessible with the current technologies. Additional directions can be found in the presentation and discussion slides on the workshop's website: <http://sites.udel.edu/uhf-nmr-workshop/2015/08/10/program/>.

Biomedical sciences

Structural, molecular, cell biology

- *Structural biology of medically and fundamentally important biomacromolecular systems:*
 - *Understanding the structure and dynamics of integral and peripheral membrane proteins (including GPCRs) in native-like environments.* Through the characterization of native

structures, the characterization of dynamics and conformational exchange, unique mechanistic insights and understanding of these highly important drug targets will be obtained. Membrane proteins represent the majority of all drug targets, that are typically undercharacterized and often mischaracterized, such as multidrug transporters targeted by antimicrobials, and receptors targeted by CNS drugs. To understand the chemistry conducted by these important proteins, they need to be studied in the context of native-like and native membranes, and these analyses will be enabled by dramatic increases in resolution and sensitivity obtained at high magnetic fields.

- Understanding the nature of heterogeneous noncrystalline biological assemblies involved in neurodegeneration and pathogenesis, such as i) amyloid fibers and oligomeric states that are critical for our understanding of Alzheimer's and related diseases; ii) multicomponent assemblies that constitute viral and bacterial pathogens and that are critical to our understanding of infectious diseases; and iii) molecular machines and motor protein assemblies in the context of intracellular trafficking that are critical for our understanding of numerous neurodegenerative disorders. These systems need the dramatically improved resolution and sensitivity from high field magnets for full characterization.
- Intrinsically disordered biomacromolecular systems, including low-population transient states involved in catalysis, molecular recognition and metabolic regulatory systems— a new frontier where only NMR is capable of characterizing nascent structure. Spectral dispersion provided by high magnetic fields is critical for this research.
- As biologics play an increasingly important role, there is a pressing need in drug development to better resolve the complex spectra of intact monoclonal antibodies and other large proteins, RNA, and their complexes, both as a means for validating manufacturing processes but also for characterizing interaction of these ligands with their large targets, both structurally and dynamically. Spectral dispersion attained at high magnetic fields is critical for the characterization of these systems.
- Pleiomorphic ensembles (membraneless cellular bodies) both *in vitro* and in intact cells. These systems are particularly challenging because of disorder (both spatial and temporal) and complexity (multiple components). The phenomenon of liquid-liquid phase separation that drives the formation of these ensembles is opening completely new areas of cell biology and biophysics. Ultrahigh magnetic fields are needed to provide resolution and sensitivity for investigating these systems.
- Biomolecules that cannot be investigated at high concentrations due to limited solubility, availability or aggregation. Many proteins and other biomolecules that have "sticky" regions fall into this category. These systems require greatly improved sensitivity and resolution

that will be afforded by high field magnets.

- Metabolomics and in-cell imaging and spectroscopy are a rapidly expanding frontier in biochemical research: the biggest payoffs are expected in this area where sensitivity and resolution directly enhance the ability to quantify metabolic changes and identify intracellular processes. Particular breakthroughs are anticipated in the field of personalized or precision medicine, where metabolite detection will be done in the human body, by integrated NMR and MRI, for diagnostics or analysis of drug effects.
- Bioinorganic chemistry and chemical biology of metallobiomolecules and pharmaceuticals through measurements of previously inaccessible quadrupolar nuclei. These are critical for studying i) studying physiological processes or for identifying metal binding sites and protonation states in biomolecules; ii) structural analysis of pharmaceutical polymorphs. The use of ultra high magnetic fields enables quadrupolar spectroscopy for many of such systems. Particularly impressive gains are expected for the measurements of rare but biologically essential nuclei, such as ^{43}Ca , ^{67}Zn , ^{25}Mg , ^{17}O , ^{33}S , $^{37/39}\text{Cl}$, ^{41}K as well as for other half-integer quadrupolar transition metals.

Physiology, Brain And Developmental Biology

- Major increases in field strengths will enable unprecedented resolution for human imaging, including
 - Laminar and columnar resolved fMRI over large regions of cortex and imaging methods for neuronal fiber tracking will enable definition of the range of normal architecture in the human brain as a basis for understanding the architectures associated with: autism, affective disorders, behavioral disorders, dyslexia, epilepsies, learning disorders, and schizophrenia.
 - Increase in resolution and sensitivity will enable functional mapping of the neuronal connectivity of the living human brain during sensory, motor, and cognitive activities.
 - Brain development in primates and anatomical development in embryology can be realized at sub-millimeter resolutions.
 - *In vivo* and sequential brain studies in concussion, Alzheimer's disease, and dementias involving protein aggregates are enabled by susceptibility contrast increases with high field.
 - Imaging of brain stem nuclei; cortical structure in normal brain and disease (MS, epilepsy); structural changes with function and dysfunction.
 - Cardiac imaging at unprecedented resolution and speed in non-human primates and other animal models.
 - New contrasts that will arise at UHF and which will enable the distinction among currently orphaned pathologies.

- Major increases in sensitivity, spectral dispersion, and relaxation parameter changes with field increases, will allow us to determine
 - Regional concentrations of the entire major bioenergetic phosphorous molecules, and the relevant biochemical kinetics evaluated using chemical exchange saturation transfer – particularly of amide protons.
 - Small molecules associated with aberrant human behavior, including serine, glycine, gamma-aminobutyric acid, dopamine and possibly folic acid. These will become amenable at less than 0.5 mM concentrations.
 - The action of Na⁺, K⁺ ATPase [NKA], the brain's most vital enzyme, will become accessible to imaging at less than 1 mm resolution in the human brain. This will be feasible by dynamic contrast enhancement and measurements of mean intracellular water molecule lifetimes, related to the enzyme's activity.
- Low-gamma and quadrupolar nuclei MRI and MRS will become feasible at high fields
 - Because of the reduced quadrupolar coupling at high magnetic fields, biological chemistry at oxygen sites rather than at carbon sites will be enabled using ¹⁷O.
 - Intracellular vs. extracellular concentrations of sodium and potassium can be evaluated on a regional basis in the living brain.
- B₀ and B₁ field distortions at high MR proton frequencies will enable imaging the electrical properties of the brain, leading to hitherto untapped sources of contrast and information.
 - Status of electrical permittivity and relations to Na and K gradients in normal and abnormal human behavior can be best evaluated at high fields/frequencies.

Chemistry, Catalysis, Materials Sciences

- Characterization of surfaces as opposed to bulk samples, enabled by the large gains in inherent sensitivity and resolution through use of ultrahigh fields. This could lead to particular improvements in characterization of catalytic surfaces, with ample implications in the chemical and energy-related industries.
- Characterization of reactive sites in catalysts, energy-related, and structural materials, enabled through the enhanced sensitivity and resolution of NMR signals from quadrupolar nuclei with non-integer-spin at high magnetic fields:
 - Solid-acid sites in aluminosilicate zeolites or silica-alumina heterogeneous catalysts to improve their activities or selectivities for processing petroleum, natural gas, or biomass-derived into diverse hydrocarbon products. Enhanced sensitivity and resolution is expected to enable examination of materials (catalysts) with minute concentrations of sites and/or much smaller surface areas than currently possible;

- Controlled hydration of aluminosilicate cementitious materials with improved compositions and structural properties under extreme conditions (deep water oil wells, geothermal wells), with reduced carbon footprints, or in support of modeling analyses to predict cement setting and properties;
- Metal coordination environments in metal oxides, such as ceramics, semiconductors, ion-conducting solids, and dielectric materials, including in thin films for device applications;
- Metal chalcogenide (sulfides, selenides) in semiconductors and opto-electronic materials;
- Complex metal hydrides for energy storage.
- Characterization of paramagnetically displaced NMR signals in catalysts, energy-related, and structural materials, enabled by enhanced resolution and sensitivity at high magnetic fields:
 - Rare-earth-doped solid-state phosphor materials for solid-state lighting applications (in conjunction with fast MAS);
 - Energy-storage materials: super-capacitors; batteries;
 - Fe-containing cementitious materials;
 - Non-precious-metal catalysts containing Ni, Co, Mn, Fe;
- Characterization of low-gamma or dilute nuclei sites in biological and inorganic materials, enabled by enhanced sensitivity and resolution at high magnetic fields:
 - ^{43}Ca NMR of biominerals, bone and teeth, and cementitious materials;
 - ^{33}S NMR studies of deactivation processes in heterogeneous catalysts, hydrodesulfurization of fossil fuels;
 - ^{15}N NMR of nitrides, such as GaN solid-state lighting materials, ceramics, or N-functionalized porous carbons for electrochemical devices;
 - ^{17}O NMR of heterogeneous catalysts, solid-oxide fuel cell materials, battery materials, and oxide semiconductors
 - ^{13}C and ^{25}Mg NMR of carbon sequestration materials;
 - Lanthanide inclusion in glasses and lasing materials;
 - Membrane materials for water purification, osmosis, desalination;
 - ^{25}Mg NMR of batteries, cements, ceramics, and metal organic frameworks (MOFs);
 - ^{239}Pu NMR of oxides and related radioactive materials;
 - ^1H -detected 2D HETCOR spectroscopy involving low-gamma nuclei: combining increased sensitivity with excellent ^1H resolution under ultrafast MAS would lead to tremendous opportunities in catalysis, biorenewables, nanocomposites, etc.
- In situ UHF NMR measurements at elevated temperatures or pressures will enable to monitor challenging material compositions and structures:
 - During hydrothermal syntheses of inorganic materials: catalysts, cements, ceramics;

- Of heterogeneous catalysts under industrially realistic operating conditions: hydrocarbon conversion, automotive emission control, etc;
- Of energy conversion and storage processes, including batteries, fuel cells, photovoltaic materials complex metal hydrides, batteries.

Biotechnology, Bioengineering, Environmental Engineering, Biogeochemistry

- *UHF will open new frontiers in bioengineering and biomaterials, including studies of:*
 - Biomineralization processes;
 - Biofilm formation, compositions, and structures;
 - Complex carbohydrates in energy-rich plant cell walls and biotechnology relevant microorganisms;
 - Membrane - bound protein - protein complexes by solution and solid - state NMR for structure-function relations of proteins at membrane interfaces, in synthetic host materials, and metabolism of drugs by enzymes;
 - Structure-function relations of proteins and inorganic species in bone matrices and their roles in bone diseases;
 - Influences of amyloid inhibitors on formation of AD brain plaques.
 - Structural studies that will help the deconstruction of lignocelluloses in plant biomasses.
- *UHF MR will also serve environmental engineering and biogeochemistry:*
 - Soil-mineral matrices;
 - Transport of dilute metal ions or organic pollutants in soils;
 - Automotive emissions catalysts.

The workshop participants stressed the fact that transformational science requires integration of NMR and MRI, to gain unprecedented new insights into systems molecular sciences. Integration with other experimental and computational techniques is also deemed to be a “must” in modern multidisciplinary research. Among the multiple examples of the synergies between NMR and MRI that were presented at the meeting, a case worth highlighting is provided by **studies of neurodegeneration**. Alzheimer’s, Parkinson’s and Lewy body dementia, traumatic brain injury, and age-related macular degeneration have in common the conversion of proteins from soluble to insoluble states. UHF solution and solid-state NMR and human brain and eye MRI can converge to yield an understanding that leads to possible treatment of these diseases. Indeed, ***NMR-based*** molecular structure studies, in solution and in the solid state, allow for the characterization of important proteins including **Intrinsically Disordered Proteins**, which are considered to be of central importance to neurodegenerative disorders. These biomolecules cannot be studied by most analytic techniques, other than NMR. The utmost sensitivity and resolution afforded by high magnetic fields

is required to overcome the narrow chemical shift dispersion and the need for low concentrations to limit aggregation. At the same time, *MRI* is uniquely poised to characterize alterations in brain neuronal architecture and inflammation associated with neurodegeneration. **UHF beyond 7 T MRI is expected to detect protein aggregates, including amyloid and tau protein, associated with dementias and brain trauma (e.g. concussion)** because of the magnetic susceptibility contrast associated with aggregates less than 100 micrometers. The same mechanism should prove useful for distinguishing these widespread neurodegenerative disorders from less common ones.

The participants stressed the unprecedented breadth of contemporary scientific directions that require MR-based UHF technologies, and the new synergies that such breakthroughs in UHFs will promote between NMR and MRI. The participants were also gravely concerned that the US was losing its competitiveness in many of the scientific areas described by the NMR science drivers and that urgent action is needed to restore the US leadership in these scientific arenas by investing in high field NMR/MRI instrumentation.

B. UHF MR: Technology development

Scientists from NHMFL (Tallahassee, FL), RIKEN (Yokohama, Japan), Jastec (Kobe, Japan), MIT (Cambridge, MA), and Bruker BioSpin (Karlsruhe, Germany) presented the recent developments of UHF magnets beyond 28.1 T. These ultrahigh magnetic fields are required to enable transformational science discussed in the previous section. To attain such fields in the NMR magnets, the participants concluded that high-temperature superconducting (HTS) materials are required. It was discussed that limitations in scientific advances can be overcome by recent developments such as the first test-coils operating in a stable superconducting state at a magnetic field ≥ 33 T [2,3] and the first demonstrations of high-stability/high-resolution multidimensional protein NMR spectra based on a non-persistent superconducting arrangement [4,5]. At the core of these two achievements is the development of new HTS materials by scientists in the US and elsewhere into practical, high-strength, long-length conductors capable of operating at fields well above 30 T. While HTS materials have been known since 1986, earlier versions had insufficient strength and/or current-density, or required impractical manufacturing processes that precluded their use in ultra-high-field magnets. Three new conductors have emerged in recent years (2007 SuperPower REBCO tape with 600 MPa strength [6], 2012 NHMFL over-pressure-processed isotropic Bi2212 [7], and 2014 Sumitomo Ni-Cr-reinforced Bi2223 [8]). The incorporation of HTS materials together with LTS materials into actively stabilized LTS/HTS duplex designs operating in non-persistent mode promise to lead to a number of important breakthroughs that can affect the long-term future of magnetic resonance. The first of these developments in ultra-high field magnets is best exemplified by the NbTi/NbSn₃/YBCO user magnet, funded by NSF (DMR 0923070). This 32-T all-superconducting magnet is being constructed by NHMFL scientists *via* the NHMFL user program. By combining five LTS coils with

two HTS coils in a serial, concentric assembly, this magnet will deliver in 2016 a 36% peak field increase over currently existing all-superconducting magnets [9]. The second breakthrough is NMR-quality test-coils using HTS materials. This was first demonstrated by the Japanese ultrahigh field NMR effort in 2009 utilizing LTS and HTS coils at 9.4 and 2.3 T, respectively, running in an ultrastabilized mode that delivered a conventional set of 500 MHz high-resolution biomolecular NMR spectra despite running non-persistently [3, 4]. In recent months the Japanese group reached 1020 MHz (23.9 T) with a similar test-coil system [10]. All these features can enable a US-based pre-eminence in HTS-based magnet technologies – a pre-eminence which so far has not translated into state-of-the-art ultrahigh field US facilities in either NMR or MRI.

Developments in HTS materials provide compelling arguments for abandoning conventional magnet preconceptions, and rising to the two grand challenges mentioned earlier by setting a roadmap for the future of ultrahigh field NMR and MRI magnets based on serial LTS/HTS magnet technologies. Given the current state of the art in HTS materials and coils, an appropriate team of experts can develop magnets with sufficient homogeneity and stability for performing NMR as well as microimaging experiments at fields in excess of 30 T and human MRI studies in fields of >14 T. Moreover, by developing the common technologies needed to construct such magnets and placing the resulting systems at the service of the NMR, microimaging and MRI communities, such a project would cut the Gordian Knot currently holding back those disciplines, which rely on higher fields for continuing progress.

A parallel approach was presented at the workshop by the Bruker Biospin scientists, who are developing ultrastable commercial magnets at 28.1 T using mostly LTS materials, and a small HTS insert. The first of such magnets is projected to be deployed in Europe in late 2017 or early 2018.

Additional discussion points are summarized below.

Magnets (NMR)

- Contemporary high-resolution NMR spectrometers are available up to ^1H frequencies of 1 GHz, based on LTS wire technologies delivering fields of up to 23.5T. LTS technologies, however, cannot be used beyond this threshold; from 23.5 T onwards, the inclusion of HTS inside the inner windings of the magnet coil becomes a necessity. This joint LTS/HTS technology is shared by the numerous development projects currently in place towards a 1.2 GHz NMR. Achievement of 28 T fields is a technically challenging and costly process, it is a feasible one. The ongoing development of an all-superconductor 32T LTS/HTS magnet with sub-ppt homogeneity at the US NHMFL, predicts that the development of a 32.8T, 1.4 GHz NMR-compatible magnet is also within reach.
- While technically feasible, it is likely that the first such UHF NMR magnet made will involve a number of compromises, of the kind that subsequent magnets will strive to minimize. Important aspects to overcome in near future design refinements include magnet size (a parameter where, as

mentioned, the LTS/HTS ratio becomes crucial), stray field size, and cryogenic consumption. In the case of a powered LTS/HTS magnet, field stability and homogeneity might need further optimization.

- Given the uncertainties mentioned above, it would certainly be productive to have a multi-group effort working independently on alternative technologies, but sharing know-hows towards the achievement of the same (or similar) UHF NMR goal. In the case of the HTS research this multiplicity of projects might be more costly at the beginning, but would eventually pay off by leading to optimized materials and technologies to be used in repeated platforms. In this respect it would be important to support three or more parallel material technologies –in particular those based on REBCO, Bi2212 and Bi2223 conductors– as well as on magnets of different quenching/insulation designs. This would help cement a solid basis for UHF NMR magnet developments for the next 10+ years, opening up the optimum strategies and timelines –not only for high resolution magnets in the 1.2/1.3 GHz range, but for NMR systems operating at 2 GHz and beyond. As part of this process, choosing the best approach for the LTS/HTS combination is the correct strategy before embarking on one or two NMR construction designs.

Magnets (MRI)

- Moving beyond 11.7 T will require adopting higher field conductors, primarily Nb₃Sn, as well as HTS-based conductors relying on either ReBCO, Bi2212 or Bi2223 superconductors. These may have to be incorporated into cables as well. A staged development program would therefore suit best an UHF MRI magnet development program, with intermediate steps in field. The challenge in pursuing the incremental steps in field with a bore size to accommodate the human head are not linear, given that the stored energy and stresses in magnets scale with square of the magnetic field.

Probes (NMR)

Maximizing the performance of the RF probeheads in solution and solid-state NMR experiments is crucial and highly dependent on the RF frequencies and the specific experiments to be performed. Probes play two critical roles in NMR. They first excite the signal by impressing multiple strong (for solids, ~20 G) and spatially very uniform RF magnetic fields over the sample. These fields are typically applied at three or four frequencies as much as a decade apart. In this process, the probe's electronic components may be required to withstand several thousand volts. A few microseconds after the strong fields are switched off, the probe then detects efficiently small NMR signals, which are then passed to a low-noise preamplifier. For RF decoupling, irradiation continues at one frequency while detection occurs at another. The probe must separate the very strong from the very weak magnetic fields. Mechanical systems in RF probeheads are also critical, in that they control the sample temperature and provide for spinning the sample. Without

optimized probes, spectroscopists cannot take full advantage of the magnetic field strength and can neither control the experimental conditions nor accurately detect the signals that arise. Representative examples of probe technologies that are envisioned to be important in the context of ultrahigh magnetic field applications include:

- **¹H solution NMR probes, including low-loss cryoprobes** - At the lower end of the anticipated field range to be generated by HTS magnets, the sensitivity is already limited by loss in conductive samples. It is currently not widely appreciated that sensitivity is limited by dielectric loss in commonly used polar organic solvents. As magnets approach 30 T, current probe designs will produce electromagnetic standing waves in standard 5-mm aqueous samples that will compromise RF field uniformity. The answer may lie in adapting the coil technologies developed for high field MRI, such as quadrature birdcage resonators and array coils, to microfabrication techniques.
- **Cryoprobes optimized for heteronuclear detection, and microsample probes for solution NMR** - High field is a great opportunity for heteronucleus detection and microsample detection, because in these cases the sensitivity is not generally limited by sample loss, and so the scaling with field strength is closer to the theoretical $B_0^{7/4}$ for spin 1/2 isotopes. Direct detection of ¹³C has advantages in mixture analysis due its large chemical shift dispersion, and is valuable for metabolomics and in particular for tracer studies. At high field, indirect detection probes experience low sensitivity with aqueous samples or even polar organic solvents. The degradation is more significant with cryogenic coils and electronics, since the loss not only reduces Q but also increases the system noise temperature. It has been shown that distorting the cross section of the sample tube from a circle into a slab or oval with its long axis parallel to the B₁ field tends to reduce ¹H loading effects and increase sensitivity. As fields increase and the shaped sample is no longer enough, it may be effective to borrow probe designs from high field MRI and to implement quadrature or even array configurations.
- **Probes optimized for detection of quadrupolar nuclei in solid-state NMR** – Ultrahigh magnetic fields will have major impact on the detection of quadrupolar nuclei, which comprise two thirds of the Periodic Table and are found in a wide variety of natural and man made materials. To realize full benefits of UHF, specialized probes designed for detection of a broad range of quadrupoles under static and moderate-frequency MAS conditions is required.
- **Ultrafast magic-angle-spinning (MAS) NMR probes** optimized for heteronuclear detection- Spinning frequencies >60 kHz is a key new capability with a broad range of applications to materials and biological systems. This requires the development of probes that enable more routine use, including sample packing, stable variable temperature (of significant range, -50 to +50 °C as a minimum), with full HX and HXY tuning ranges.

- **Probes optimized for ^1H detection in solid-state NMR** - A major opportunity is high-resolution ^1H -detected solid state NMR of large biomolecules. This technique requires ultra-high magnetic fields and ultrafast (>100 kHz) sample rotation. Considerable progress has already been made on micro-MAS spinners, but continued development to enhance stability and improve sample handling is required.
- **Low temperature probes for LN_2 and < 80 K for MAS NMR** - Many MAS spectra of proteins and nucleic acids are missing lines in the spectra due to “dynamic processes” which interfere with the ^1H decoupling. Thus, low temperature spectroscopy provides access to these regions of proteins. At the moment low temperature probes operating to 100 K are commercially available. Operating at lower temperature provides opportunities for higher sensitivity because of larger Boltzmann factors and also DNP. Two groups have recently described closed cycle recirculation systems for these experiments. It is important that they be made commercially available.
- **High-frequency and DNP-compatible oriented-samples probes for membrane research** - A strategy to use routinely DNP with oriented samples and availability of commercial probes would be a game changer, particularly at UHF, where dramatic sensitivity and resolution gains are anticipated.
- **Microimaging UHF NMR probes for characterizations of energy-storage materials (*in situ* batteries, super-capacitors, water filtration polymers)** - Ultrahigh magnetic fields will enable detection of signals from quadrupolar nuclei and those displaced paramagnetically in energy-storage materials, through dramatically enhanced sensitivity and resolution. To realize full potential for *in situ* characterization of these materials, microimaging probes are required to visualize structural changes in a spatially- and temporally-resolved manner.
- **High-frequency sources for NMR** - As NMR moves into the microwave “L band” it is possible to take advantage of many advances in electronics technology developed for military and communications purposes. In particular, the continued improvement in high frequency RF power transistors utilizing GaAs and more recently GaN technology is improving the performance (the noise figure and linearity) and reducing the cost and size of ^1H NMR power amplifiers in the 1 – 2 GHz range.

Microwave sources for DNP NMR at ≥ 1.2 GHz

Dynamic nuclear polarization (DNP) is an increasingly fundamental component of solid-state NMR research, which has particular promise as the discipline moves into UHF research. In particular, with the recent discovery of Overhauser effects in insulators it appears that DNP will function more efficiently at higher fields. The three possible microwave sources for DNP at ≥ 1200 MHz / 789 GHz consists of (1) diodes, (2) extended interaction klystrons (EIKs) operating as oscillators and (3) gyrotrons also operating as

oscillators. These platforms should be further investigated; so should be the development of gyro-amplifiers so that pulsed DNP experiments can be performed at these fields.

- At the moment diodes do not have sufficient power above 95 GHz to be useful for DNP except possibly at ~4.2 K and using a high Q resonator.
- Currently EIK's are appearing that operate at 263 GHz / 400 MHz but because of the small size of the slow wave structure they currently do not operate at higher 395 GHz / 600 MHz or higher. However, as technology evolves this situation may change.
- Thus, at the moment the only source that can supply >20 watts at the ~789 GHz needed for 1.2 GHz NMR investigations, is the gyrotron. Presently oscillators are operating in second harmonic mode at 527 GHz. The challenge would be to develop an oscillator operating in 3rd harmonic to reduce the cost of adding DNP to an existing system.
- Pulsed DNP experiments in principle circumvent the ω_0^{-1} or ω_0^{-2} dependence of the cross and solid effects, respectively. Thus, the development of amplifiers operating is an exciting new area for the development of DNP.

Gradients (MRI)

- MRI is currently running at or near the physiological limit imposed by peripheral nerve stimulation threshold limits; a major problem will be to significantly speed up EPI encoding in light of the biological limits. Still, the developments of ultrastrong gradients associated with the Human Connectome project can provide solutions to this limitation.

RF Rx Technology (MRI)

- Parallel receiver (Rx) technology is a relatively well understood problem. Arrays around between 64 and 128 channels are practical and doable without major hurdles.

RF Tx Technology (MRI)

- While initially considered a major hurdle at UHF's, it is now agreed that parallel transmit (Tx) can provide good coverage and uniform excitation –at least in brain and musculoskeletal studies– well beyond 7 T. We need more analysis of high channel Tx arrays for brain MRI at >14T. Parallel transmit will likely be mandatory for good excitation homogeneity. Preliminary studies suggest conventional pTx will work. Body heating temperature constraints will dictate for pulse sequence designs and useful coil configurations.
- Much useful work can be done with low-SAR sequences.
- On-coil amplification may be important as an enabling technology; safety issues will be one of the major hurdles for some of the applications.

Pulse sequences for the quantitative mapping of small molecules in the human brain

- UHF magnetic resonance spectroscopy *in vivo* has the potentials for detection and mapping of molecules involved in brain energetic as well as essential neuroreceptor pathways and bioamine pathways. This detection of crucial metabolites will be facilitated both when targeting them by direct spectroscopic (MRS) avenues, as well as by indirect CEST/MT-based measurements benefiting from the UHFs.

Motion mitigation/correction (MRI)

- As UHFs will enable functional and anatomic studies approaching 0.1 mm resolutions, patient motion could become a large issue. We need improved real-time feedback or other approaches.

Acquisition strategies which minimize the biological hemodynamic spatial response in fMRI studies

- As fields increase and spatial resolution improves, new functional MRI opportunities will arise that exceed the localization limits of brain activation currently provided by hemodynamic fMRI means (BOLD, CBF, CBV etc.). Questions arose as what will be the new contrast fMRI mechanisms arising at >14T fields, and what will be their resolution limits. Judging from pre-clinical studies, it is likely that new contrast agents going beyond the T_2^* window opened by the BOLD effect will open, including diffusion, T_2 , T_1 and perfusion mechanisms changing the observable MR signal.
- Equally important will be the search for new functional markers of cognitive activities, emerging as the fields increase substantially past 7T.

Biological effects of UHF MRI

- While detrimental biological effects associated with the exposure of humans to fields beyond 9.4 T have not been observed, studies of animal and human exposure to fields in excess of 12 T will need to be undertaken to assess the fate of the 14 T project. This can be carried out at facilities with available magnets tailored for relatively large animal studies (dogs, pigs); low-homogeneity ≥ 13 T field facilities also exist, where humans could be accommodated and the effects of static and/or slowly changing magnetic fields be assessed.
- Additional considerations to be assessed in UHF human studies relate to safety and physiological effects associated by pulsing gradients and RF at the high fields and high frequencies to be involved.

Additional information can be found in the presentation and discussion slides on the workshop's website: <http://sites.udel.edu/uhf-nmr-workshop/2015/08/10/program/>.

C. UHF MR: Instrument deployment and funding mechanisms, shared facilities infrastructure

It was the consensus of the Workshop's participants that a successful implementation of UHF MR technologies for enabling next generation transformational science will require establishing sustainable shared centers that will house UHF instrumentation and which will provide support for users, personnel, instrumentation, and physical infrastructure. At the workshop, the participants discussed extensively the requirements and models for such shared facilities. Two leaders of European ultrahigh field NMR centers, Drs. Lucia Banci and Lyndon Emsley, shared the highly successful models that are implemented by the European Union in such facilities. The directors of shared UHF NMR facilities in the US outlined the current models used in their centers. At the brainstorming discussion sessions current challenges and key prerequisites for the development of sustainable UHF infrastructure were outlined. These are summarized below.

Models for shared UHF facilities in the US

The establishment is needed for a sufficient number of National High Field NMR & MRI Facilities that develop enabling leading edge technologies in a broad range of science, to take optimal advantage of ultrahigh magnetic fields and make them available to broad scientific communities throughout the US.

- ✓ These facilities will have state of the art commercial instrumentation as well as the ability to develop instrumentation for specific needs at the cutting edge of spectroscopy, imaging and science. There is a great deal of room for these developments especially in an environment in which there is only one major NMR spectrometer manufacturer. In addition to probes, field stabilization technology and novel field shimming capabilities may be required, the development of field gradient technology is also going to be necessary.
- ✓ A single portal for access to the high field NMR/MRI facilities is suggested so that researchers can choose the best facility for their research needs.
- ✓ These facilities will serve the entire high field NMR spectroscopy community in the US, much of it through remote access. While remote access works quite well today (as exemplified especially by PNNL, NHMFL, and NMRFAM facilities) there is a need for enhancing this capability. In order to serve the broad community a well-functioning website must be developed that provides critical information and facilitates access.

- ✓ Each facility does not have to be optimized for all user activities. While many activities will likely be in common, others such as ultra high temperature for materials would be established at only one or two facilities until the demand merits expansion into other facilities. Different focus in each center is expected as well as complementary expertise, with some overlaps in capabilities and scientific themes.
- ✓ Shared protocols, equipment and reagents will lead to improved efficient usage of the instruments. The community has demonstrated their interest and willingness to cooperate in this way.
- ✓ These facilities will establish close working relationships with other facilities (such as APS) to leverage complementary technologies, e.g., SAXS, crystallography, cryo-EM, computation, multimodal imaging. This will be accomplished by identifying infrastructure opportunities for integrating NMR/MRI/MRS science with other complementary methodologies.

While the establishment of National High Field MR Facilities is the vision for housing the next generations of high field magnets, there is an urgent need for current and immediate next generation instruments for the biological, chemical and biomedical communities in the United States. An interim solution must be found so that American scientists can continue to pursue cutting edge materials, chemical, and biological science. This includes the need for 1 and 1.2 GHz as well as DNP instruments housed in local or regional facilities.

Infrastructure- mechanisms for sustained support

Models need to be established for continued and sustainable funding mechanisms to support:

- ✓ Innovative infrastructure, equipment and highly qualified in-house personnel
- ✓ Equipment acquisition, operation, and maintenance in these facilities.
- ✓ Research and training of in-house and off-campus investigators.
- ✓ Off-campus investigator access (travel and accommodations) and usage of the facilities.

The participants recognized the critical importance of personnel development. To this end, it is essential to:

- ✓ Establish mechanisms for attracting and sustaining young investigators for high field technology and applications. Restore competitiveness of the US UHF MR science on the international arena.
- ✓ Establish mechanisms to recruit and retain trained technical and scientific personnel of the highest caliber in shared resources; this will ensure that users and on-site scientists will mutually push technological advances forward.

It was stressed that likely sustainable support models will include partnerships between Federal, State, industry, and philanthropic organizations.

Additional information can be found in the presentation and discussion slides, on the workshop's website:
<http://sites.udel.edu/uhf-nmr-workshop/2015/08/10/program/>.

The participants concluded that in-depth analysis needs to be conducted by working groups, to identify the viable sustainable model(s) for shared UHF centers. The results of the analysis will be presented in the roadmap document.

III. SUMMARY AND NEXT STEPS

The US scientific community has an urgent need of UHF MR technologies and infrastructure that are required to address a broad range of societally important problems of strategic national interest. This UHF infrastructure is currently lacking in the US, leading to a progressive loss of the US's scientific community leadership position in a number of scientific arenas on the international scene. To address the pressing scientific needs and develop sustainable UHF MR infrastructure for the nation, we recommend that the following steps be immediately adopted:

1. A roadmap document be compiled through this post-workshop effort and backed by the scientific community at large. This document will contain detailed analyses of the key scientific, technological, and infrastructure development issues mentioned in Section II of this report, followed by suggestions on how to exploit emerging UHF for the sake of maximizing US scientific and technological leadership. For the general outline of strategies planned to develop the roadmap see **Appendix A**.
2. The final roadmap be discussed and developed with the various agencies that hosted and participated in this workshop. Involvement of other organizations, such as AAAS, national and international professional societies, and private foundations will be encouraged.

IV. REFERENCES CITED

- [1] "High Magnetic Field Science and Its Application in the United States: Current Status and Future Directions". Committee to Assess the Current Status and Future Direction of High Magnetic Field Science in the United States; Board on Physics and Astronomy; Division on Engineering and Physical Sciences; National Research Council. 232 pages, ISBN: 978-0-309-28634-3. (2014)
- [2] Markiewicz, et al., "33.8 Tesla with a $\text{YBa}_2\text{Cu}_3\text{O}_{7-x}$ superconducting test coil", *Adv. Cryog. Eng.*, 55 (6), 225-230 (2010).
- [3] Trociewitz, U.P.; Dalban-Canassy, M.; Hannion, M.; Hilton, D.K.; Jaroszynski, J.; Noyes, P.; Viouchkov, Y.; Weijers, H.W. and Larbalestier, D.C., 35.4 T field generated using a layer-wound superconducting coil made of $(\text{RE})\text{Ba}_2\text{Cu}_3\text{O}_{7-x}$ (RE = rare earth) coated conductor, *Appl. Phys. Lett.*, 99 (20), 202506 (2011).
- [4] Yanagisawa, Y.; Nakagome, H.; Hosono, M.; Hamada, M.; Kiyoshi, T.; Hobo, F.; Takahashi, M.; Yamazaki, T.; and Maeda, H., Towards beyond-1 GHz solution NMR: Internal ^2H lock operation in an external current mode, *J. Magn. Reson.* 192, 329–337 (2008).
- [5] Kiyoshi, T.; Choi, S.; Matsumoto, S.; Zaitzu, K.; Hase, T.; Miyazaki, T.; Otsuka, A.; Yoshikawa, M.; Hamada, M.; Hosono, M.; Yanagisawa, Y.; Nakagome, H.; Takahashi, M.; Yamazaki, T.; Maeda, H., HTS-NMR: Present Status and Future Plan *IEEE Trans. App. Supercon.* 20, 714-717 (2010).
- [6] Majkic, G.; Galstyan, E.; Selvamanickam, V., High Performance 2G-HTS Wire Using a Novel MOCVD System, *IEEE Tran. On Applied Supercond.*, 25 (3), 6605304 (2015).
- [7] Larbalestier, D. C.; Jiang, J.; Trociewitz, U. P.; Kametani, F.; Scheuerlein, C.; Dalban-Canassy, M.; Matras, M.; Chen, P.; Craig, N. C.; Lee, P. J.; Hellstrom, E. E., Isotropic round-wire multifilament cuprate superconductor for generation of magnetic fields above 30 T, *Nature Materials* 13, 375-381 (2014).
- [8] Nakashima, T.; Yamazaki, K.; Kobayashi, S.; Kagiya, T.; Kikuchi, M.; Takeda, S.; Osabe, G.; Fujikami, J.; Osamura, J., Drastic Improvement in Mechanical Properties of DI-BSCCO Wire with Novel Lamination Material", *IEEE Tran. On Applied Supercond.*, 25 (3), 6400705 (2015).
- [9] Weijers, H.W.; Markiewicz, W.D.; Voran, A.J.; Gundlach, S.R.; Sheppard, W.R.; Jarvis, B.; Johnson, Z.L.; Noyes, P.D.; Lu, J.; Kandel, H.; Bai, H.; Gavrilin, A.V.; Viouchkov, Y.L.; Larbalestier, D.C. and Abraimov, D.V., Progress in the Development of a Superconducting 32 T Magnet with REBCO High Field Coils, *IEEE Trans. Appl. Supercond.*, 24, 5 (2014).
- [10] Nishiyama, Y.; Pandey, M. K.; Florian, P.; Fyon, F.; Hashi, K.; Ohki, S.; Nishijima, G.; Matsumoto, S.; Noguchi, T.; Deguchi, T.; Goto, A.; Shimizu, T.; Maeda, H.; Takahashi, M.; Yanagisawa, Y.; Tanaka, R.; Nemoto, T.; Miyamoto, T.; Suematsu, H.; Saito, K.; Miki, T., 1020 MHz LTS/HTS NMR: II. Application to solid-state NMR. Presented at the 56th Experimental Nuclear Magnetic Resonance Conference (ENC), Asilomar, CA, April 19-24, 2015.

APPENDIX A

UHF MR SCIENCES AND INFRASTRUCTURE IN THE USA: TOWARD A ROADMAP FOR THE COMING DECADE

The proposed strategy for development of a detailed roadmap consists of forming working groups as outlined below. The composition, leadership, format and timeline for these groups will be developed after we receive feedback from the scientific community regarding proposals developed during this workshop, and after discussions with the agencies whose missions are relevant to the workshop goals.

I. SCIENCE NEEDS AND BENEFITS

1. Working group on science drivers for biomedical sciences
 - a. NMR
 - b. MRI
 - c. Integrated MR
2. Working group on science drivers for materials, catalysis, energy sciences
 - a. NMR
 - b. MRI
 - c. Integrated MR

II. TECHNOLOGY DEVELOPMENT

1. Working group on high-temperature superconductor wires and cables
2. Working group on superconducting magnet design and installation
3. Working group on MRI-specific hardware components: fast and high amplitude field gradients, shimming, parallel RF transceiver coils
4. Working group on NMR-specific spectrometer hardware: cryogenic probes for solution and solids, ultrafast magic-angle-spinning, microimaging, terahertz sources for DNP
5. Working group on physiological effects and safety of MRI and MRS for human studies at magnetic fields >9.4 T

III. INSTRUMENT DEPLOYMENT AND FUNDING MECHANISMS: SHARED FACILITIES AND INFRASTRUCTURE

1. Working group on sustainable models for shared UHF MR centers in the US
 - a. Models for shared centers
 - b. Cost analysis and infrastructure requirements
 - c. Funding models
2. Working group on advancing UHF MR-based science in the US
 - a. Support of the broader scientific community
 - b. Analysis of socioeconomic benefits
 - c. Public relations (website, advertising materials, videos etc.)

IV. SUMMARY RECOMMENDATIONS

APPENDIX B

MEETING PARTICIPANT LIST

Assa-Munt, Nuria

National Institutes of Health, Center for Scientific Review
Biological Chemistry and Macromolecular Biophysics
6701 Rockledge Drive, MSC 7806
Bethesda, MD 20817
United States
assamunu@csr.nih.gov

Banci, Lucia

University of Florence
CERM & Department of Chemistry
Via Luigi Sacconi 6
Sesto Fiorentino, 50019
Italy
banci@cerm.unifi.it

Baum, Jean

Rutgers, The State University of New Jersey
Department of Chemistry and Chemical Biology
610 Taylor Road
Piscataway, NJ 8854
United States
baum@chem.rutgers.edu

Bax, Ad

NIDDK, National Institutes of Health
Bldg 5, Rm 126, 9000 Rockville Pike
Bethesda, MD 20892
United States
bax@nih.gov

Bayer, Paul

U.S. Department of Energy
Office of Biological and Environmental Research
U.S. Department of Energy, SC-23.1, Germantown Bldg,
1000 Independence Ave., SW
Washington, DC 20585
United States
paul.bayer@science.doe.gov

Begg, Michael

Tesla Engineering Limited
Unit 1, Water Lane
Storrington, RH20 3EA
United Kingdom
begg@tesla.co.uk

Bird, Mark

Florida State University
MagLab
1800 E Paul Dirac Dr
Tallahassee, FL 32306
United States
bird@magnet.fsu.edu

Boebinger, Gregory

National High Magnetic Field Laboratory
Director's Office
1800 East Dirac Drive
Tallahassee, FL 32310
United States
gsb@magnet.fsu.edu

Budinger, Thomas

University of California, Berkeley
Bioengineering
Bldg. 55. Lawrence Berkeley Natl. Lab. Berkeley, CA
Berkeley, CA 94720
United States
tfbudinger@lbl.gov

Byrd, R. Andrew

National Cancer Institute
Structural Biophysics Laboratory
538 Chandler St
Frederick, MD, MD 21702-1201
United States
byrdra@mail.nih.gov

Chekmenov, Eduard

Vanderbilt University
Institute of Imaging Science / Department of Radiology
and Radiological Sciences
708 Winsley Place
Brentwood, TN 37027
United States
eduard.chekmenov@vanderbilt.edu

Chin, Jean

National Institutes of Health, NIGMS
Division of Cell Biology and Biophysics
Bethesda, MD 20892-6200
United States
chinj@nigms.nih.gov

Chmelka, Brad

University of California Santa Barbara
Dept. of Chemical Engineering
University of California, Santa Barbara, Dept. of Chemical
Engineering, Mail Code 5080
Santa Barbara, CA 93106
United States
bradc@engineering.ucsb.edu

MEETING PARTICIPANT LIST

Cross, Timothy

Florida State University
Chemistry and Biochemistry
1800 E. Paul Dirac Dr.
Tallahassee, FL 32310
United States
cross@magnet.fsu.edu

Doty, David

Doty Scientific
R&D
700 Clemson Rd
Columbia, SC 29229
United States
david@dotynmr.com

Duyn, Jeff

NINDS, National Institutes of Health
9000 Rockville Pike, Bldg. 10, Room B1D724
Bethesda, MD 20892-1065
United States
jhd@helix.nih.gov

Emsley, Lyndon

Ecole Polytechnique Fédérale de Lausanne (EPFL)
Chemistry
Avenue Forel
Lausanne, 1015
Switzerland
lyndon.emsley@epfl.ch

Fiechtner, Gregory

DOE/Basic Energy Sciences
Chemical Sciences, Geosciences & Biosciences Division
19901 Germantown Road
Germantown, MD 20874-1920
United States
gregory.fiechtner@science.doe.gov

Frydman, Lucio

Weizmann Institute
Chemical Physics Department
Rehovot, 76100
Israel
lucio.frydman@weizmann.ac.il

Ghose, Ranajeet

National Science Foundation
Molecular Biophysics
4201 Wilson Blvd
Arlington, VA 22230
United States
rgghose@nsf.gov

Glover, Kerney Jebrell

Lehigh University
Chemistry 6 E.
Bethlehem, PA 18015
United States
kjd206@lehigh.edu

Grant, Steven

National Institute on Drug Abuse NIH
Division of Clinical Neuroscience and Behavioral Research
6001 Executive Blvd Room 3278
Bethesda, MD 20892
United States
sgrant@nida.nih.gov

Griffin, Robert G.

Massachusetts Institute of Technology
Francis Bitter Magnet Lab
170 Albany Street, NW14
Cambridge, MA 02139
United States
rgg@mit.edu

Gronenborn, Angela M

University of Pittsburgh, School of Medicine
Structural Biology
1051 BST3, 3501 Fifth Ave
Pittsburgh, PA 15260
United States
amg100@pitt.edu

Hamada, Mamoru

Japan Superconductor Technology, Inc.
Administration
1-5-5 Takatsuka-dai, Nishi-ku
Kobe, 651-2271
Japan
hamada.mamoru@kobelco.com

MEETING PARTICIPANT LIST

Hayes, Sophia

Washington University
Chemistry
1 Brookings Dr., #1134
St. Louis, MO 63130
United States
hayes@wustl.edu

Hinks, R. Scott

GE Healthcare
Applied Science Laboratory
W226 N174 Takoma Dr
Waukesha, WI 53186
United States
scott.hinks@med.ge.com

Hoch, Jeffrey

UConn Health
Molecular Biology and Biophysics
263 Farmington Ave.
Farmington, CT 06030-3305
United States
hoch@uchc.edu

Hong, Mei

Massachusetts Institute of Technology
Department of Chemistry
170 Albany Street, NW14-3212
Cambridge, MA 2139
United States
meihong@mit.edu

Igumenova, Tatyana

Texas A&M University
Biochemistry and Biophysics
300 Olsen Boulevard
College Station, TX 77843
United States
tigumenova@tamu.edu

Ishii, Yoshitaka

University of Illinois at Chicago
Chemistry
845 W Taylor St, SES Rm4500
Chicago, IL 60607
United States
yishii@uic.edu

Iwasa, Yukikazu

Massachusetts Institute of Technology
Francis Bitter Magnet Laboratory, Plasma Science and
Fusion Center
NW14-3101 170 Albany Street
Cambridge, MA 2139
United States
iwasa@jokaku.mit.edu

Jaroniec, Christopher

The Ohio State University
Chemistry and Biochemistry
222 CBEC Building, 151 West Woodruff Ave.
Columbus, OH 43210
United States
jaroniec.1@osu.edu

Jerschow, Alexej

New York University
Department of Chemistry
100 Washington Square East
New York, NY 10012
United States
alexej.jerschow@nyu.edu

Klosek, Malgorzata

NIH
ORIP
6701 Democracy Blvd
Bethesda, MD 20892
United States
klosekm@mail.nih.gov

Larbalestier, David

Florida State University
Applied Superconductivity Center
2031 East Paul Dirac Drive
Tallahassee, FL 32310
United States
larbalestier@asc.magnet.fsu.edu

Laukien, Frank

Bruker Corporation
40 Manning Road
Billerica, MA 1821
United States
Frank.Laukien@bruker.com

MEETING PARTICIPANT LIST

Levy, Abraham

National Institutes of Health
Office of the Director
6701 Democracy Blvd
Bethesda, MD 20892
United States
levyabra@mail.nih.gov

Long, Joanna

University of Florida
Biochemistry & Molecular Biology
Box 100245
Gainesville, FL 32610-0245
United States
jrlong@ufl.edu

Maas, Werner

Bruker BioSpin Corp
MRS
15 Fortune Drive
Billerica, MA 1821
United States
werner.maas@bruker.com

Maeda, Hideaki

RIKEN
NMR Facility
1-7-22 Suehiro-cho, Tsurumi-ku
Yokohama, 230-0045
Japan
maeda@jota.gsc.riken.jp

Markley, John

University of Wisconsin-Madison
Biochemistry
433 Babcock Drive
Madison, WI 53706
United States
jmarkley@wisc.edu

McDermott, Ann

Columbia University
Chemistry
MC 3113
New York, NY 10027
United States
aem5@columbia.edu

Melhem, Ziad

Oxford Instruments
Alliances
Oxford Instruments, Tubney Woods
Abingdon, OX13 5QX
United Kingdom
ziad.melhem@oxinst.com

Miklos, Andrew

National Institutes of Health
National Institute of General Medical Sciences
45 Center
Bethesda, MD 20892-6200
United States
andrew.miklos@nih.gov

Minervini, Joseph

Massachusetts Institute of Technology
Plasma Science and Fusion Center
77 Massachusetts Avenue, NW22-129
Cambridge, MA 2139
United States
minervini@psfc.mit.edu

Mueller, Karl

Pacific Northwest National Laboratory
Physical & Computational Sciences
902 Battelle Blvd K9-80
Richland, WA 99352
United States
karl.mueller@pnnl.gov

Murillo, Carlos

National Science Foundation (NSF)
Chemistry Division
4201 Wilson Blvd, Suite 1054
Arlington, VA 22230
United States
cmurillo@nsf.gov

Niendorf, Thoralf

Max-Delbrueck Center for Molecular Medicine
Berlin Ultrahigh Field Facility (B.U.F.F.)
Robert Roessle Strasse 10
Berlin, 13125
Germany
thoralf.niendorf@mdc-berlin.de

MEETING PARTICIPANT LIST

Nishiyama, Yusuke

JEOL RESONANCE Inc.
Development group
3-1-2 Musashino, Akishima
Tokyo, 196-8558
Japan
yunishiy@jeol.co.jp

Opella, Stanley

University of California, San Diego
Chemistry and Biochemistry
9500 Gilman Drive 307
La Jolla, CA 92093-0307
United States
sopella@ucsd.edu

Palmer, Arthur

Columbia University
Biochemistry and Molecular Biophysics
Hammer 616, Box 36; 701 West 168th Street
New York, NY 10025
United States
agp6@columbia.edu

Pederson, Mark

Basic Energy Sciences
Chemical Sciences, Geosciences, and Biological Sciences
19901 Germantown Road, Room E-433
Germantown, MD 20874
United States
mark.pederson@science.doe.gov

Perl, Daniel

Uniformed Services University of the Health Sciences
Pathology
4301 Jones Bridge Road, Room B-3138
Bethesda, MD 20814
United States
daniel.perl@usuhs.edu

Pettigrew, Roderic

National Institute of Biomedical Imaging and
Bioengineering
Office of the Director
31 Center Drive
Bethesda, MD 20814
United States
roderic.pettigrew@nih.gov

Pietrass, Tanja

DOE
BES/CSGB
SC 22.1/Germantown Building, 1000 Independence Ave.,
SW
Washington, DC 20585
United States
tanja.pietrass@science.doe.gov

Polenova, Tatyana

University of Delaware
Chemistry and Biochemistry
036 Brown Laboratories
Newark, DE 19716
United States
tpolenov@udel.edu

Preusch, Peter

National Institute of General Medical Sciences
Division of Cell Biology and Biophysics
45 Center Drive
Bethesda, MD 20892
United States
preuschp@nigms.nih.gov

Pruski, Marek

Ames Laboratory, Iowa State University
Chemistry
230 Spedding Hall
Ames, IA 50011
United States
mpruski@iastate.edu

Ramamoorthy, Ayyalusamy

University of Michigan
Chemistry
Chemistry, 930 N. University Ave, University of Michigan
Ann Arbor, MI 48109-1055
United States
ramamoor@umich.edu

Rienstra, Chad

University of Illinois Urbana Champaign
Chemistry
A129 CLSL box 50-6 M/C 712 600 S Mathews Ave.
Urbana, IL 61801
United States
rienstra@illinois.edu

MEETING PARTICIPANT LIST

Rosen, Bruce

Massachusetts General Hospital
Radiology
Bldg. 149, 13th Street, Room 2301D
Charlestown, MA 02129
United States
bruce@nmr.mgh.harvard.edu

Rutt, Brian

Stanford School of Medicine
Radiology
1201 Welch Road
Stanford, CA 94305
United States
brutt@stanford.edu

Roth, Gerhard

Bruker Biospin
Magnets
Wikingerstr. 13
Karlsruhe 76189
Germany
gerhard.roth@bruker.com

Schaaf, Michael

Siemens Healthcare GmbH
HC ID MR TR
Allee am Roethelheimpark 2
Erlangen 91052
Germany
michael.schaaf@siemens.com

Sherry, A. Dean

UTSouthwestern Medical Center
Advanced Imaging Research Center
5323 Harry Hines Blvd. NE4.210
Dallas, TX 75390-8568
United States
dean.sherry@utsouthwestern.edu

Sheely, Douglas

National Institutes of Health, NIGMS
45 Center Drive
Bethesda, MD 20892
United States
douglas.sheeley@nih.gov

Sodickson, Daniel

New York University School of Medicine
Radiology
660 First Avenue, Fourth Floor
New York, NY 10016
United States
Daniel.Sodickson@med.nyu.edu

Spinu, Leonard

National Science Foundation
Division of Materials Research
National Science Foundation, 4201 Wilson Blvd
Arlington, VA 22230
United States
LSpinu@nsf.gov

Springer, Charles

Oregon Health & Science University
Advanced Imaging Research Center
3165 S.W. 70th Avenue
Portland, OR 97225
United States
springer@ohsu.edu

Strauss, Bruce

U S Department of Energy
Office of High Energy Physics
1000 Independence Ave SW
Washington, DC 20585-1290
United States
bruce.strauss@science.doe.gov

Stringer, John

PhoenixNMR
Owner
510 E. 5th Street
Loveland, CO 80537
United States
js@phoenixnmr.com

Swain, Amy

National Institutes of Health, NIGMS
Division of Biomedical Technology, Bioinformatics, and
Computational Biology
45 Center Drive, MSC 6200
Bethesda, MD 20892-6200
United States
SwainA@mail.nih.gov

Tessema, Guebre X.

National Science Foundation
Division of Materials Research
4201 Wilson Blvd
Arlington, VA 22230
United States
gtessema@nsf.gov

MEETING PARTICIPANT LIST

Thulborn, Keith

University of Illinois
Center for Magnetic Resonance Research
1745 Telegraph Road
Bannockburn, IL 60015
United States
mrix@ameritech.net

Tolbert, Blanton

Case Western Reserve University
Chemistry
10900 Euclid Ave
Cleveland, OH 44106
United States
bst18@case.edu

Tycko, Robert

National Institutes of Health
Laboratory of Chemical Physics, NIDDK
Building 5, Room 112
Bethesda, MD 20892-0520
United States
robertty@mail.nih.gov de

Ugurbil, Kamil

University of Minnesota
Center for Magnetic Resonance Research
2021 6th Street SE
Minneapolis, MN 55455
United States
kamil@cmrr.umn.edu

Wagner, Gerhard

Harvard Medical School
Biological Chemistry and Molecular Pharmacology
240 Longwood Avenue
Boston, MA 2115
United States
gerhard_wagner@hms.harvard.edu

Wald, Lawrence

Massachusetts General Hospital-Harvard Medical School
A. A. Martinos Center for Biomedical Imaging
Bldg 149 13th St rm 2301
Charlestown, MA 2129
United States
wald@nmr.mgh.harvard.edu

Wehrle, Janna

National Institute of General Medical Sciences, NIH
Division of Cell Biology and Biophysics
45 Center Drive, Rm 2AS.13B
Bethesda, MD 20892-6200
United States
wehrlej@nigms.nih.gov

Wright, Peter

The Scripps Research Institute
Integrative Structural and Computational Biology
10550 North Torrey Pines Road, MB204
La Jolla, CA 92037
United States
wright@scripps.edu

Zilm, Kurt

Yale University
Chemistry
350 Edwards Street
New Haven, CT 6511
United States
kurt.zilm@yale.edu

APPENDIX C

PROGRAM

THURSDAY, NOVEMBER 12, 2015

(attendees arrive at the Visitor entrance at 07:30 for NIH entry)

08:00–08:20	Lister Lobby	Arrival and Registration
08:20–08:30	Lister Auditorium	Introduction and Welcome Remarks Tatyana Polenova, Thomas Budinger
08:30–10:00	Lister Auditorium	Plenary Session 1A: <i>Low- and High-Temperature Superconducting Magnet Technologies in MRI and NMR</i> Session Chair: Roderic Pettigrew, Director, NIBIB 8:30–8:50 State of the Art and Challenges in Ultrahigh Field Magnet Technology (David LARBALÉSTIER, NHMFL) 8:50–9:10 Toward Super-High Field and Ultra-Compact Size NMR Magnets Operated Beyond 1 GHz (Hideaki Maeda, RIKEN) 9:10–9:30 UHF Magnet Development at MIT (Yukikazu Iwasa, Francis Bitter Magnet Laboratory, MIT) Plenary Session 1B: <i>Shared Ultrahigh Field NMR Facilities in Europe</i> Session Chair: Robert Tycko, NIDDK 9:30–9:45 European Large Scale Facilities for NMR Spectroscopy (Lyndon Emsley, École Polytechnique Fédérale de Lausanne) 9:45–10:00 NMR Infrastructures in Europe (Lucia Banci, CERM, University of Florence)
10:00–10:30	Lister Lobby	Coffee Break
10:30–12:15	Lister Auditorium	Parallel Session 2A: <i>Frontiers of NMR</i> Session Chair: Lucio Frydman, National High Magnetic Field Laboratory, The Weizmann Institute of Science 10:30–10:55 Emerging Trends in Solution NMR (Gerhard Wagner, Harvard University) 10:55–11:05 Discussion 11:05–11:25 Emerging Trends in Biosolids NMR (Ann McDermott, Columbia University) 11:25–11:35 – Discussion 11:35–12:05 Emerging Trends in NMR of Materials (Marek Pruski, Ames Laboratory) 12:05 – 12:15 - Discussion
10:30–12:15	The Claude Denson Pepper Building (31), floor 6, room 5	Parallel Session 2B: <i>Frontiers of MRI</i> Session Chair: Jeff Duyn, NINDS 10:30 –10:55 Functional Neuroimaging at Ultrahigh Fields (Bruce Rosen, Harvard Medical School and MIT) 10:55 – 11:05 Discussion 11:05 – 11:25 Non-Proton Metabolic Imaging and Anatomical Proton Imaging (Keith Thulborn, University of Illinois Chicago) 11:25 – 11:35 Discussion 11:35 – 12:05 Emerging Contrasts At Ultrahigh Fields (A. Dean Sherry, University of Texas Dallas, UT Southwestern Medical Center) 12:05 – 12:15 - Discussion
12:15–1:30	NIH cafeteria	Lunch

PROGRAM

1:30-3:15	Lister Auditorium	Plenary Session 3: Frontiers in Ultrahigh Field Magnet Technologies Session Chair: Chad Rienstra, University of Illinois at Urbana-Champaign 1:30-1:55 Development of 1020 MHz NMR Superconducting Magnet Using Bi-2223 Innermost Coil (Mamoru Hamada, Jastec) 1:55-2:20 - UHF Magnets at NHMFL (Mark Bird, NHMFL) 2:20-2:45 - UHF Magnets at Bruker (Gerhard Roth, Bruker) 2:45-3:10 - Designing Ultrahigh Field Magnets for NMR and MRI (Joseph Minervini, MIT)
3:15-3:45	Lister Lobby	Coffee Break
3:45-6:00	Lister Auditorium	Round Table Discussion Session 4: Science Drivers for Ultrahigh Field NMR Science in the US Discussion Leader: Angela M. Gronenborn, University of Pittsburgh Scribe: Jean Baum, Rutgers University
7:00-9:30	Redwood Restaurant and Bar	Session 5: Dinner Followed by After Dinner Address Frontiers in Magnetic Resonance:- Current Perspective and Looking into the Future (Greg Boebinger, Director, NHMFL) Session Chair: Brad Chmelka, UC Santa Barbara

PROGRAM

FRIDAY, NOVEMBER 13, 2015

08:30–9:40	Lister Auditorium	Breakout Discussion Session 6A: <i>Future Challenges and Barriers for NMR Science in the US</i> Discussion Leader: Jeff Hoch, University of Connecticut Scribe: Mei Hong, MIT
08:30–9:40	The Claude Denson Pepper Building (31), floor 6, room 5	Breakout Discussion Session 6B: <i>Future Challenges and Barriers for MRI Science in the US</i> Discussion Leader: Kamil Ugurbil, University of Minnesota Scribe: Eduard Chekmenev, Vanderbilt University
9:40–10:10	Lister Auditorium	Breakout Session 7A: <i>Shared Ultrahigh Field NMR Facilities in the US</i> Session Chair: Joanna Long, University of Florida Scribe: Christopher Jaroniec, The Ohio State University 9:40–9:45 National High Magnetic Field Laboratory (Tim Cross, NHMFL) 9:45–9:50 The National Magnetic Resonance Facility at Madison (John Markley, University of Wisconsin) 9:50–9:55 Environmental Molecular Sciences Laboratory (EMSL) (Karl Mueller, Pacific Northwest National Laboratory) 9:55–10:00 MIT-Harvard Center for Magnetic Resonance (Robert Griffin, MIT) 10:00–10:05 Biotechnology Resource Center for NMR Molecular Imaging of Proteins (Stanley Opella, University of California San Diego) 10:05–10:10 New York Structural Biology Center (Arthur G. Palmer III, Columbia University)
9:40–10:10	The Claude Denson Pepper Building (31), floor 6, room 5	Breakout Session 7B: <i>Ultrahigh Field Magnet Development for MRI</i> Session Chair: Joseph Minervini 9:40–9:45 UHF Magnet Development at Oxford Instruments (Ziad Melhem) 9:45–9:50 UHF Magnet Development at Tesla (Michael Begg) 9:50–9:55 UHF MRI at Siemens (Michael Schaaf) 9:55–10:00 UHF MRI at GE (Scott Hinks) 10:00–10:10 Discussion
10:10–10:30	Lister Lobby/ The Claude Denson Pepper Building (31), floor 6, room 5	Coffee Break

10:30–12:15	Lister Auditorium	<p>Parallel Session 8A: <i>Emerging NMR Technologies in Liquids and Solids</i> Session Chair: Kurt Zilm, Yale University 10:30–10:45 Biomolecular Solution NMR at Ultrahigh Fields (Ad Bax, NIH) 10:45–10:50 Discussion 10:50–11:05 Solution NMR Studies of Intrinsically Unstructured Proteins at Ultrahigh Fields (Peter Wright, The Scripps Research Institute) 11:05–11:10 Discussion 11:10–11:30 Challenges of ultrahigh field DNP NMR (Robert Griffin, MIT) 11:30–11:35 Discussion 11:35–11:50 MAS NMR Probes for Ultrafast MAS: challenges for Ultrahigh Fields (John Stringer, PhoenixNMR) 11:50–11:55 Discussion 11:55–12:10 NMR of Half-Integer Quadrupolar Nuclei in Materials: Opportunities at Ultrahigh Fields (Sophia Hayes, Washington University) 12:10–12:15 Discussion</p>
10:30–12:15	The Claude Denson Pepper Building (31), floor 6, room 5	<p>Parallel Session 8B: <i>Emerging MRI/MRS Technologies</i> Session Chair: Alexej Jerschow, New York University 10:30–10:45 Ultrahigh Field MRI and Potential Clinical Breakthroughs (Thoralf Niendorf, Max-Delbrueck Center for Molecular Medicine) 10:45–10:50 Discussion 10:50–11:05 High Field Imaging Gradients and Coils (Lawrence Wald, Massachusetts General Hospital-Harvard Medical School) 11:05–11:10 Discussion 11:10–11:30 The Active Transport of Water Molecules in Biological Tissues: Underpinnings of MRI Interpretation (Charles Springer, Oregon Health Sciences University) 11:30–11:35 Discussion 11:35–11:50 <i>In Vivo</i> MRI/MRS and Hyperpolarization: Opportunities and Challenges at UHF (Eduard Chekmenev, Vanderbilt University)- CANCELED 11:50–11:55 Discussion 11:55–12:10 Opportunities for MRI/MRS/Pathology Correlations in the Human Brain (Daniel Perl, Uniform Services University) 12:10–12:15 Discussion</p>
12:15–1:30	NIH campus cafeteria	Lunch
1:30–3:15	Lister Auditorium	<p>Round Table Discussion Session 9: <i>Development of LTS and HTS Magnet Technologies in MRI and NMR: Roadmap</i> Discussion Leader: Thomas Budinger, UC Berkeley Scribe: Tatyana Polenova, University of Delaware</p>
3:15–3:45	Lister Lobby	Coffee Break
3:45–6:00	Lister Auditorium	<p>Round Table Discussion Session 10: <i>Development of a roadmap for ultrahigh field NMR/MRI/MRS science in the US</i> Discussion Leader: Tatyana Polenova, University of Delaware Scribe: Thomas Budinger, University of California Berkeley</p>
6:00		Adjourn

PRESENTATION TITLES (Abstracts for these presentations are posted on the workshop's website)

SESSION 1A: Low- and High-Temperature Superconducting Magnet Technologies in MRI and NMR

State of the Art and Challenges in Ultrahigh Field Magnet Technology (David Larbalestier, NHMFL)
Toward Super-High Field and Ultra-Compact Size NMR Magnets Operated Beyond 1 GHz (Hideaki Maeda, RIKEN)
UHF Magnet Development at MIT (Yukikazu Iwasa, Francis Bitter Magnet Laboratory, MIT)

SESSION 1B: Shared Ultrahigh Field NMR Facilities in Europe

European Large Scale Facilities for NMR Spectroscopy (Lyndon Emsley, École Polytechnique Fédérale de Lausanne)
NMR Infrastructures in Europe (Lucia Banci, CERM, University of Florence)

SESSION 2A: Frontiers of NMR

Emerging Trends in Solution NMR (Gerhard Wagner, Harvard University)
Emerging Trends in Biosolids NMR (Ann McDermott, Columbia University)
Emerging Trends in NMR of Materials (Marek Pruski, Ames Laboratory)

SESSION 2B: Frontiers of MRI

Functional Neuroimaging at Ultrahigh Fields (Bruce Rosen, Harvard Medical School and MIT)
Non-Proton Metabolic Imaging and Anatomical Proton Imaging (Keith Thulborn, University of Illinois Chicago)
Emerging Contrasts At Ultrahigh Fields (A. Dean Sherry, University of Texas Dallas, UT Southwestern Medical Center)

SESSION 3: Frontiers in Ultrahigh Field Magnet Technologies

Development of 1020 MHz NMR Superconducting Magnet Using Bi-2223 Innermost Coil (Mamoru Hamada, Jastec)
UHF Magnets at NHMFL (Mark Bird, NHMFL)
UHF Magnets at Bruker (Gerhard Roth, Bruker)
Designing Ultrahigh Field Magnets for NMR and MRI (Joseph Minervini, MIT)

SESSION 5: Dinner Followed by After Dinner Address

Frontiers in Magnetic Resonance:- Current Perspective and Looking into the Future (Greg Boebinger, Director, NHMFL)

SESSION 7A: Shared Ultrahigh Field NMR Facilities in the US

National High Magnetic Field Laboratory (Tim Cross, NHMFL)
The National Magnetic Resonance Facility at Madison (John Markley, University of Wisconsin)
Environmental Molecular Sciences Laboratory (EMSL) (Karl Mueller, Pacific Northwest National Laboratory)
MIT-Harvard Center for Magnetic Resonance (Robert Griffin, MIT)
Biotechnology Resource Center for NMR Molecular Imaging of Proteins (Stanley Opella, University of California San Diego)
New York Structural Biology Center (Arthur G. Palmer III, Columbia University)

SESSION 7B: Ultrahigh Field Magnet Development for MRI

UHF Magnet Development at Oxford Instruments (Ziad Melhem)
UHF Magnet Development at Tesla (Michael Begg)
UHF MRI at Siemens (Michael Schaaf)
UHF MRI at GE (Scott Hinks)

SESSION 8A: Emerging NMR Technologies in Liquids and Solids

Biomolecular Solution NMR at Ultrahigh Fields (Ad Bax, NIH)
Solution NMR Studies of Intrinsically Unstructured Proteins at Ultrahigh Fields (Peter Wright, The Scripps Research Institute)
Challenges of ultrahigh field DNP NMR (Robert Griffin, MIT)
MAS NMR Probes for Ultrafast MAS: challenges for Ultrahigh Fields (John Stringer, PhoenixNMR)
NMR of Half-Integer Quadrupolar Nuclei in Materials: Opportunities at Ultrahigh Fields (Sophia Hayes, Washington University)

SESSION 8B: Emerging MRI/MRS Technologies

Ultrahigh Field MRI and Potential Clinical Breakthroughs (Thoralf Niendorf, Max-Delbrueck Center for Molecular Medicine)
High Field Imaging Gradients and Coils (Lawrence Wald, Massachusetts General Hospital-Harvard Medical School)
The Active Transport of Water Molecules in Biological Tissues: Underpinnings of MRI Interpretation (Charles Springer, Oregon Health Sciences University)
Opportunities for MRI/MRS/Pathology Correlations in the Human Brain (Daniel Perl, Uniform Services University)